

Hemoglobin A1c

Analyzer



HbA1c Monitoring System:

- No pretreatment
- No pipetting
- Small Sample Volume
- All-in-one Analyzer
- Fully automatic operation (3 steps)

ANALYZER SPECIFICATION	DESCRIPTION
System Description	Point-of-care- immunoassay analyzer
Measuring Method	Photometry
Sample Volume	Less than 0.2 μ L
Sample Preparation	No pretreatment; no pipetting required
Sample Type	Capillary whole blood from fingertips or anticoagulated venous whole blood (EDTA)
Quantitative test	Hemoglobin A1c (whole blood) Range: 4% ~16%
Time	HbA1c < 8 mins
Display Unit	Percentage (%), mg/dL, or mmol/mol
Precision	CV <5%
Patient ID	Barcode reader or keyboard (Optional)
Memory	10,000 results
CARTRIDGE SPECIFICATION	
Unit per Carton	20
Storage	2-8 °C (35.6 - 46.4 °F)
Dimensions	5.51 (H) x 5.51 (W) x 12.8 (D) inch 14 (H) x 14 (W) x 32.5 (D) cm



FHA1c.01